

# MICRO SCALE PURIFICATION SYSTEMS FOR BIOLOGICAL SAMPLE PREPARATION

Thayne Edwards<sup>1</sup>, Bruce K. Gale<sup>2</sup>, A. Bruno Frazier<sup>1</sup>  
School of Electrical & Computer Engineering, Georgia Institute of Technology  
777 Atlantic Drive, Atlanta GA 30332-0250  
E-mail: [Bruno.Frazier@ece.gatech.edu](mailto:Bruno.Frazier@ece.gatech.edu)

Department of Biomedical Engineering, Louisiana Tech University  
Ruston, LA 27338

## ABSTRACT

This work is focused on the development of miniaturized instrumentation formats for integrated biochemical and biological sample preparation. In particular, the paper is focused on micro fluidic systems for the purification of cells, uncoated nano particles, and surface coated nano particles. Micro fluidic systems will be described for purifying cells from whole blood, sorting of blood cells based on their electrophysiological characteristics, and separating and purifying the components of a blood cell homogenate. Additionally, the micro sample preparation systems will be shown to be effective in purifying complex samples of nano particles based on the particles physical size and effective electrical charge. The micro systems will be demonstrated for the purification of samples containing polystyrene nano particles with multiple diameters as well as for samples containing a mixture of uncoated and surface coated nano particles. The formats discussed in this work include the micro electrophysiological characterization system; the micro thermal field flow fractionation system; and the micro electrical field flow fractionation system. The micro cellular electrophysiological characterization system is used for cell sorting and selection. The cell sorting and selection is accomplished using impedance spectroscopy on single cells. The system has been shown to be effective in sorting blood cells by cell type (i.e., red, white) and by sub populations (e.g. viable leukocytes, non viable leukocytes). The micro thermal field flow fractionation system is a chromatographic separation technique for fractionating particle samples based on the heat capacity, thermal conductivity and physical size of the particles. The micro electrical field flow fractionation system fractionates particle samples based on the physical size and zeta potential (effective electrical charge) of the particles within the sample. Electrical and thermal micro FFF systems are capable of separating particle samples with constituents in the diameter range from approximately 1 nm to 1  $\mu\text{m}$ . The separation and purification of a variety of nano particles will be demonstrated in the field flow fractionation systems including cells, cellular sub-components, uncoated polystyrene nano particles, and protein coated (Protein A) polystyrene nano particles.

## INTRODUCTION

The miniaturization of biochemical analysis systems has been a topic of growing interest over the past decade. During this time, a large majority of the technical effort has been invested in the development of the primary separation (or amplification) component of the various analysis systems such as the micro columns used in the miniaturized chromatographic systems (e.g. electrophoresis, gas chromatography, liquid chromatography) or the chambers used for miniaturized polymerase chain reaction (PCR) systems. Less effort have been directed toward the development of technologies for micro scale sample preparation (with the exception of PCR). Sample preparation technologies include methods for purifying, manipulating, interfacing, amplifying, and

chemically modifying sub-micro liter volumes of samples for analysis in a miniaturized format. While each of these technologies is available in a macro scale format, most have not been available on the micro scale until recently. During the past two to three years, there has been a significant increase in the worldwide efforts to improve the technology base for miniaturized sample preparation. These efforts have resulted in micro systems for purifying and sorting biological samples, manipulating samples, sample amplification, and interfacing samples with micro analysis systems [1].

In this paper, we will focus on the use of micro systems fabrication technologies for the development of three different micro systems for use in integrated sample preparation. The three systems include: a) the micro electrical field flow fractionation system; b) the micro thermal field flow fractionation system; and c) micro electrophysiological characterization system.

### **MICRO ELECTROPHYSIOLOGICAL CHARACTERIZATION ( $\mu$ -EPC)**

Electric impedance (EI) measured in a one-dimensional electric field is an established method for interrogating the electromagnetic behavior of isolated materials and composite systems. Micro fabrication techniques offer a low-power means of applying traditional EI concepts to the investigation of complex-valued dielectric properties of small structures and material samples. Reducing the overall size of an EI measurement device allows for increased spatial resolution while limiting the possibility of dielectric breakdown by minimizing the strength of the required electromagnetic field. In addition, micro scale EI devices require only nano liters of sample solution.

Micro fabricated EI sensors have demonstrated the ability to recognize variations in solution temperatures, ionic concentrations, hydrogen peroxide concentrations, and even antigen-antibody binding (immuno-sensors) [2,3]. Previously reported systems interrogate solutions using an array of surface mounted metal electrodes with an active surface area of  $\sim 1 \text{ mm}^2$ . To further reduce the required sample size we fabricated EI systems containing micro channels lined with metal electrodes. The devices allow materials to be positioned between interrogating electrodes using rapid and efficient fluid transport methods [4-6]. Although micro scale EI devices may have numerous applications in basic science and engineering, their size makes them especially suitable for biological studies.

Investigations into the dielectric properties of proteins is one of many possible applications for an integrated micro channel/EI measurement system. The electrical characteristics of protein solutions are believed to play an important role in their physiologic functions which involve protein-protein and charged ligand interactions [7-9]. In these studies, the dielectric dispersion of protein solutions is attributed to a frequency dependence on permittivity and attempts were made to relate the observed electrical characteristics to the summation of distributed charges within the protein (polarization vector) and proton fluctuations.

An additional biological application for a micro scale EI device is the measurement and extraction of the dielectric properties of single living cells. Numerous techniques have been developed to extract the canonical parameters (i.e., membrane capacitance, membrane resistance, and cytoplasmic resistance) of both individual cells and cell aggregates. The three most common techniques include the EI cell suspension technique [10-12], whole-cell patch clamping [13-15], and electrophoresis/electrorotation

[16-27]. All of these methods rely on estimates of cell geometry for the accurate calculation of the electrical parameters and are generally limited to whole-cell resolution. Micro channels outfitted with EI measurement electrodes may provide increased spatial resolution (by using electrodes smaller than a single cell) and known cellular geometry (by constraining the cell within the micro channel) [28].

The micro system is suitable for impedance measurements of femto liters of fluids, solutions, suspended particles, and single cells within micro channels. Figure 1 is a schematic of the micro system in which two fluid reservoirs are connected by a single micro channel which narrows to a width of  $\sim 10\mu\text{m}$  (channels are  $\sim 4\mu\text{m}$  high). The micro channels were constructed from epoxy-based photoresist on quartz glass wafers. Full-depth gold measurement electrodes were integrated into the narrowest portion of the micro channels. Through holes were wet etched in microscope coverslips and bonded to the microstructures to form the top surface of the micro channels. The resulting micro channels and electrodes are sandwiched between planar glass substrates exhibiting excellent optical qualities and allowing for direct sample observation using transmission light microscopy. Parasitic capacitance between the electrodes is minimized by isolating the electrodes on all sides with materials with high dielectric constants (glass and epoxy based photoresist). The micro-EI measurement system was characterized using biological concentrations of ionic salt solutions, air, and partially DI water over the frequency range of 100 Hz to 2 MHz [29].




Figure 1. Schematic view of the micro electrophysiological characterization system.

Figure 2 is a scanning electron micrograph of a micro channel with a bonded coverslip overlaid on an image of a electrode/micro channel device prior to bonding. The narrow region of the micro channel in the figure measures  $\sim 10 \mu\text{m}$  wide and  $4.3 \mu\text{m}$  deep with the gold electrodes measuring  $8.0 \mu\text{m}$  wide and  $\sim 4.0 \mu\text{m}$  thick. The gold electrodes typically extended into the channel 1 to  $2 \mu\text{m}$  on each side resulting in an average electrode gap of  $7.13 \mu\text{m}$ . Prior to coverslip bonding, the depth of the channels was measured using a Dektak IIA profilometer (Sloan Technologies, Santa Barbara, CA) and found to vary by only  $0.11 \mu\text{m}$  with a mean value of  $4.31 \mu\text{m}$ . Devices with a total of

eight different micro channel and electrode geometries were designed and fabricated on each wafer. Square holes were wet etched completely through the 200  $\mu\text{m}$  thick coverslip glass in addition to etching each coverslip free from the adjacent neighbors. The gold bonding pads are shown in Figure 2 extending from the edges of the bonded glass coverslip.




Figure 2. A SEM image of a coverslips bonded to the bottom substrate can be seen overlaid on an image of a electrode/micro channel device prior to bonding.

The  $\mu$ -EAS was used to measure the total electrical impedance (EI; designated “Z”) between the gold microelectrodes over a frequency range of 100 Hz to 2 MHz. The impedance is a function of the geometry and properties of the device as well as the interaction with the cell / media in the recording zone. To characterize the device and the electrode-electrolyte interface, we measured Z after filling with several isotropic materials having known dielectric properties. Example results are provided in Figure 3 (A,B) for air (thick solid curves), partially deionized water (dot-dashed curves) and several concentrations of phosphate buffer solution (PBS) as indicated in the figure legend. The equipment did not allow for clean records above 100 MOhm. It is notable that the micro system behaves nearly as an ideal capacitor when filled with air, as expected, but has a much more complex behavior when filled with physiological concentrations of PBS. Characteristics of the metal-electrolyte interface are primary determinants underlying the PBS impedance curves.

We have also measured electrical impedance between the microelectrodes when individual polymorphonuclear leukocytes (PMNs) and individual red blood cells (RBCs) were positioned directly between the electrodes. Example results are shown in Figure 3

(C,D) for PMNs (n=28) and in (E,F) for RBCs (n=50). Both the magnitude and phase were highly repeatable between different cells. Results for PMNs were very interesting in that two distinct subpopulations were easily identified on the basis of the EI data. Averages for type I cells are shown in thick dashed lines and averages for type II cells are shown as thick solid lines in Figure 3 (C,D). Type I cells correspond to non-viable PMNs and type II cells correspond to viable, healthy PMNs. It is well known that the permeability of cell membranes decrease upon cell death. Also, notable is the difference in the data between the PMNs and the RBCs. Clearly, the  $\mu$ -EAS was capable of distinguishing between the two cell types as well as between viable and non-viable PMNs. In either case, these preliminary studies indicate potential for the  $\mu$ -EAS in cell sorting and characterization. It is also notable that the largest intercellular differences were recorded at the highest frequency tested, 2 MHz. This is expected due to the dielectric constant of the membrane allowing a stronger trans-membrane electric field at high frequencies. Differences between cells are most easily seen in the phase for PMNs, Figure 3D, which may reflect differences in cytoarchitecture between the recording electrodes.




Figure 3. The electrical impedance was measured across the gold micro electrodes using a HP4194 Impedance Analyzer. Curves on the left (A,B) were recorded when the channel was filled with air, partially DI water and concentrations of PBS. Curves in the center (C,D) are for the same  $\mu$ -EAS when PMNs were positioned directly between the electrodes. Gray solid curves indicate individual cells, the black dashed curves show the average for PMN type I cells, and the black solid curves show the average for PMN type II cells. Results for RBCs are shown on the far right (E,F). Thick black solid curves are averages for the RBC population. Thin black solid and dashed curves are reproduced from C and D for comparison.

### MICROMACHINED THERMAL FIELD-FLOW FRACTIONATION ( $\mu$ -TFFF)

TFFF is an elution separation technique similar to chromatography except the separation field is normal to the sample and carrier flow. TFFF utilizes thermal diffusion as the separation field instead of the gel, liquid, or column packing found in

chromatographic separations. This field is accomplished by a temperature gradient across the channel. A schematic of the TFFF system is shown in Figure 4. Separation of the suspended particles is performed in a solvent carrier such as methanol, THF, acetonitrile, or DMSO. Many solvents have been used and there have been studies shown how these solvents affect separation characteristics [30]. Water is not typically used as a carrier fluid unless an electrolyte has been added. The particles in the solvent react to the temperature gradient by diffusing toward the cold wall. Higher molecular weight particles react more to the thermal gradient and are compacted more tightly against the cold wall than do lower molecular weight particles. Because of the laminar velocity profile of the carrier, samples that compact less will have a higher average velocity than the samples that compact more. The difference in average velocity results in the spatial and temporal particle separation at the output of the TFFF channel. [31]




Figure 4. Schematic of a thermal field-flow fractionation (TFFF) channel.

The TFFF system has some unique characteristics making it more suitable for some separations than some conventional systems. In TFFF, the separation field is applied normal to the separation. The resolution requirement in the direction of separation is no longer needed, which means lower field strength, lower power consumption demands, and shorter separation times. TFFF also has the advantage of elution techniques, in that the samples are collected in fractions at given times. As a result, very pure samples can be obtained. [31]

In order for a separation in a TFFF channel to occur, there must be: 1. a difference in molecular weight or diameter of the samples; 2. a sample selective perturbation of the samples toward one wall; and 3. a laminar velocity profile that results in a different average velocity of each constituent of the sample [32].

The channel height is a critical parameter is the spatial resolution of two eluting samples. The channel height is inversely proportional to this resolution. Current systems have channel widths of about  $127\mu\text{m}$  [32]. Micromachining technologies allow accurate and precise definition of channels of very small size. The channels fabricated for the  $\mu$ -TFFF device are approximately  $25\mu\text{m}$ . Smaller channels are achievable. As a result, higher resolution separations can be performed which means faster separations and more pure samples.

#### *TFFF Experimental Methods*

The  $\mu$ -TFFF system was fabricated with an n-type, [100], single side polished, silicon wafer using common micromachining processes. A silicon dioxide ( $\text{SiO}_2$ ) mask was thermally grown at  $1200^\circ\text{C}$  to approximately  $1\mu\text{m}$  thick on the surface. Photolithographic techniques were used to define the input/output ports in the photoresist on the backside of the wafer. The  $\text{SiO}_2$  mask was etched using a buffered HF solution. The ports were then anisotropically etched through the silicon wafer in 10% KOH at  $90^\circ\text{C}$ .

The SiO<sub>2</sub> was left on the wafer as an electrical insulating layer for the heaters. Titanium (Ti) was sputtered onto the backside of the wafer to about 1000 Å and patterned to define the resistive heater.

A negative photosensitive epoxy (SU-8) was spun on the wafer to a thickness of 25 μm. The channel was defined in the SU-8 by UV exposure and developing. Heat treatment of the channel walls ensured a strong bond and durability when exposed to the carrier fluid, which is typically an organic solvent. A glass microscope slide was adhered to the top of the SU-8 to complete the channel. After completing the fabrication of the μ-TFFF separation channel, the complete system was assembled (e.g. fluid interconnections, power supply, flow meters, detector, and fraction collector) and tested.

Using the μ-TFFF device fabricated as described in this paper, the total plate height was determined as a function of flow velocity for an unretained sample. DI water was used as the carrier and the sample was pure acetone. The flow rate of water was set by the pump. The flow rates used were 2.0, 1.75, 1.5, 1.25, and 1.0 mL/hr. An acetone sample, 0.2 μL, was injected into the input port at time zero for each flow rate. A detector measured the absorbance with respect to a sample of DI water.

The retention of polystyrene (PS) spheres was then tested in the system, since separation is based on the difference in retention for different sized particles. Although water is not typically used for TFFF separations because of the difficulty in performing separations in this medium, DI water was used to demonstrate the ability of the μ-TFFF system. The flow rate was set to 1.5 mL/hr to minimize the plate height. A temperature gradient of 40°C (T<sub>hot wall</sub>=70°C, T<sub>cold wall</sub>=30°C) was set up using the integrated heater and an external heat sink. A 0.2 μL sample of PS spheres (394 nm) was injected into the system.

### Micro TFFF Results




Figure 5. Profile of one end of μ-TFFF. The power is more than 1000 times lower than that reported for a typical macro-scale TFFF channel. Plate height was determined as a function of average carrier velocity and compared with results from the μ-EFFF system [33] and

One end of a fabricated channel and port is shown in Figure 5. The channel height has been reduced from 127 μm (macro-TFFF system) [7] to 50 μm (μ-TFFF). A power of only 10 W was required to achieve a 20°C temperature difference across the channel. This




Figure 6. Plate height as a function of flow velocity. This result is typical of such a channel.

TFFF theoretical results. The plate height curve is shown in Figure 6. The  $\mu$ -TFFF plate height characteristics follow the curves that are found in the  $\mu$ -EFFF channel as well as the theoretical curve for the TFFF system. The data are of lower magnitude than the sum of plate heights reported for retained polystyrene samples ( $MW_n$ :154k, 392k, and 735k).




Figure 7. Retention of PS, 394 nm spheres and separation with and unknown size PS impurity.

The temperature gradient was  $30^{\circ}\text{C}$  and toluene is the carrier. The channel dimension is  $0.025\text{cm} \times 1.2\text{cm} \times 305\text{cm}$ . [34]

The PS-394 sample retention results are found in Figure 7. Three peaks appeared as a result of the test. The first peak is the void peak which contains unretained impurities. The sample was apparently not a pure 394nm PS sample. The other peak that showed up was an impurity of unknown size. This result indicated the effectiveness of the  $\mu$ -TFFF system in purifying samples in preparation for further processing.

## MICROMACHINED ELECTRICAL FIELD- FLOW FRACTIONATION

Another system with significant potential in the area of sample preparation with which we have been working is the micromachined electrical field- flow fractionation system ( $\mu$ -EFFF).




Figure 8. Schematic diagram of the operating principle for the EFFF system.

Electrical Field-Flow Fractionation is a particle separation technique that relies on an electric field perpendicular to the direction of flow and separation as shown in Figure 8. The separations are performed in a low-viscosity liquid (typically an aqueous buffer solution) which is pumped through the separation channel. EFFF controls the relative velocity of particles by forcing particles towards the wall of the channel. Particles with high charge density pack closer to the wall and move more slowly compared to particles of lower charge density that form a more diffuse cloud and move more quickly through the channel.

The channels for the miniaturized EFFF system are shown schematically in Figure 9 and are fabricated by bonding a silicon substrate and a glass substrate together around a photolithographically defined polyimide spacer. Both substrates have metal thin films on their surface, which are patterned to define the electrodes for the channel and an




Figure 9. Schematic of channel and detector layout (a) Side view of channel (b) Top view of channel

by particle size and charge, can produce a monodisperse sample for later analysis using another system. EFFF has an additional advantage in that it is a very gentle separation method and is suitable for cells, liposomes, micelles, fragile proteins, and other delicate structures and polymers. Particles ranging in size from about 5 nm up to about 1  $\mu\text{m}$  in diameter are separable using a micromachined EFFF ( $\mu$ -EFFF) system. Combining EFFF on a chip with other analysis systems would provide the opportunity for




Figure 10. Separations of 44, 130, and 204 nm polystyrene particles with on-chip AC mode detector showing steadily increasing diameter of eluting particles. Voltage was 1.6 V, flow rate was 0.3 mL/hr and current was 27  $\mu\text{A}$ .

electrical impedance detector. The input and output ports are fabricated in the silicon substrate using KOH etching and are 200  $\mu\text{m}$  square on the interior of the channel and about 1 mm square on the external face of the silicon substrate. The channel dimensions are typically about 6 cm long, 1-6 mm wide, and 10-50  $\mu\text{m}$  in height. Fabrication and other information about the system have been reported previously [35]. The system also has an integrated electrical impedance detector that has been incorporated at the exit end of the channel [1,36,37].

One of the main advantages associated with EFFF is that it is an elution method that allows for the collection of fractions at the exit to the system. Thus, EFFF, which separates by particle size and charge, can produce a monodisperse sample for later analysis using another system. EFFF has an additional advantage in that it is a very gentle separation method and is suitable for cells, liposomes, micelles, fragile proteins, and other delicate structures and polymers. Particles ranging in size from about 5 nm up to about 1  $\mu\text{m}$  in diameter are separable using a micromachined EFFF ( $\mu$ -EFFF) system. Combining EFFF on a chip with other analysis systems would provide the opportunity for performance of a sample purification or separation step using EFFF and then, by simply redirecting the flow at appropriate intervals, allowing the analysis to continue in an orthogonal direction, using another system, or in a parallel direction using a similar system.

### Micro EFFF Results

The  $\mu$ -EFFF system has been demonstrated with a range of polymers as well as a few biological samples of interest. A typical separation is shown in Figure 10. In this figure, a high-speed separation of a three component mixture of polystyrene samples was accomplished. As demonstrated here, EFFF separates particles of similar




Figure 11. Fractograms of particles with the same diameter, but differing levels of carboxylation. The run labeled “low” had a lower density of COOH groups on the surface and a correspondingly lower surface charge, so the particles eluted sooner than did those with a greater surface charge, labeled by “high”.




Figure 12. Fractograms of showing differential retention between bare particles and particles with attached proteins.

particles with attached proteins indicates that there is a variation across the peak that corresponds to the amount of protein adsorbed to the particle. Thus, if particles were required to have a minimum level of protein attached to the particle, particles towards the end of the elution peak could be collected, or if a uniform number of attached proteins was required, a thin slice of the peak could be collected to create a monodisperse sample for later analysis.

EFFF also has the ability to retain a wide variety of particles contained in a single sample, such as a blood sample. Blood analysis is one of the most common medical

material based entirely on the particle size. Thus a polydisperse sample would show up as a wide peak on the x-axis with each slice in time or volume representing a monodisperse sample. Accordingly, collection of a small volume of sample at a given location provides a well-defined, monodisperse sample for further analysis.

The micromachined EFFF system also has the ability to separate identically sized particles based entirely on electrophoretic mobility. In this case, the particles of lower charge will elute before particles with a higher charge as shown in Figure 11. The  $\mu$ -EFFF would then also be available to prepare samples that differ only by surface charge characteristics.

A biologically valuable separation of this type was demonstrated in the  $\mu$ -EFFF system when particles with adsorbed proteins were differentially retained from those without any adsorbed protein. The results of such an analysis are presented in Figure 12. An analysis of this type is very useful in biocompatibility and protein adsorption studies. Since particles with and without protein are retained differentially, particles with adsorbed proteins could be collected for later analysis, assuming they were the particles of interest. In addition, the wide peak for the

procedures and the retention of a blood sample is demonstrated in Figure 13. While the cells found in blood are much too large to be significantly retained in a  $\mu$ -EFFF system, the proteins and other particles found in blood are quite easily retained. Thus, the cells could be separated from the smaller particles, or the smaller particles can be collected for later analysis. The same blood sample is shown for two different runs. In the first run, the retention of the whole blood is shown. The second run shows a sample that was homogenized in an ultrasonic bath. The homogenized sample shows the retention of the cell contents that were not present in the whole blood sample. In both cases, after the elution time is determined for the particle type of interest a fraction of that sample could be collected for later analysis. Thus,  $\mu$ -EFFF demonstrates its ability to function as a sample preparation system for a wide range of analytes and particles of interest.




Figure 13. Comparison of whole blood and homogenized blood retained in the  $\mu$ -EFFF system focusing on the whole blood fractogram. Notice there is continual elution of particles during most of the run, but a strong peak appears near the end of the run in a range where no elution in the homogenized sample occurred.

## CONCLUSION

In this work, three micro system formats are discussed for miniaturized biochemical sample preparation. These systems, when used in series with complimentary micro analysis systems, offer to potential to realize a total analysis system capable of sample in / answer out operation.

## ACKNOWLEDGEMENTS

Support of this work by the following sponsors is greatly appreciated: National Institute of Environmental Health Sciences ES-10846-01, Genicon Sciences Inc., MicroChem Corp.

## REFERENCES

- [1] Proceedings of the Micro-TAS 2000 Conference, Enschede, The Netherlands, May 14-18, 2000.
- [2] P. V. Gerwen, W. Laureys, G. Huyberechts, M. O. D. Beeck, and K. Baert, "Nanoscaled interdigitated electrode arrays for biochemical sensors," presented at Transducers '97, Chicago, IL, 1997.
- [3] G. R. Langereis, W. Olthuis, and P. Bergveld, "Measuring conductivity, temperature and hydrogen peroxide concentration using a single sensor structure," presented at Transducers '97, Chicago, IL, 1997.
- [4] A. Manz and H. Becker, "Parallel Capillaries for High Throughput in Electrophoretic Separations and Electroosmotic Drug Discovery Systems," presented at Transducers '97, Chicago, IL, 1997.
- [5] J. M. Ramsey, A. G. Hadd, and S. C. Jacobson, "Applications of Precise Fluid Control on Microchips," presented at Transducers '97, Chicago, IL, 1997.
- [6] L. Bousse, A. Kopf-Sill, and J. W. Parce, "An Electrophoretic Serial to Parallel Converter," presented at Transducers '97, Chicago, IL, 1997.
- [7] C. T. O'Konski, "Electric properties of macromolecules. V. Theory of ionic polarization in polyelectrolytes.," *J. Phys. Chem.*, vol. 64, pp. 605-619, 1960.

- 
- [8] D. J. Barlow and J. M. Thornton, "The distribution of charged groups in proteins," *Biopolymers*, vol. 25, pp. 1717-1733, 1986.
- [9] S. Takashima and K. Asami, "Calculation and measurement of the dipole moment of small proteins: use of protein data base.," *Biopolymers*, vol. 33, pp. 59-68, 1993.
- [10] T. Hanai, et al., "Dielectric theory of concentrated suspensions of shell-spheres in particular reference to the analysis of biological cell suspensions," *Bull. Inst. Chem. Res.*, v. 57, pp. 297-305, 1979.
- [11] H. P. Schwan, "Electrical prop. of tissue & cell suspensions," *Adv. Med. Biol. Phys.*, 147-209, 1957.
- [12] H. P. Schwan, *Determination of Biological Impedance*, vol. 6. New York: Academic Press, 1963.
- [13] I. Tasaki, *Physiology and Electrochemistry of Nerve Fibers*. New York: Academic Press, 1982.
- [14] S. Takashima, K. Asami, and Y. Takahashi, "Frequency domain studies of impedance characteristics of biological cells using micropipette technique.," *Biophys. J.*, vol. 54, pp. 995-1000, 1988.
- [15] B. Hill and D. T. Campbell, "An improved vaseline gap voltage clamp for skeletal muscle fibers," *J. Gen. Physiol.*, vol. 67, pp. 265-293, 1976.
- [16] W. M. Arnold and U. Zimmerman, "Rotating-field-induced rotation and measurement of the membrane capacitance of single mesophyll cells of *Avena sativa*," *Z. Naturforsch.*, v.37C, pp.908-915, 1982.
- [17] V. P. Pastushenko, P. I. Kuzmin, and Y. A. Chizmadshv, "Dielectrophoresis and electrorotation: a unified theory of spherically symmetrical cells.," *Stud. Biophys.*, v. 110, pp. 51-57, 1985.
- [18] R. Georgiewa, E. Donath, and R. Glaser, "On the determination of human erythrocyte intracellular conductivity by means of electrorotation - influence of osmotic pressure," *Stud. Biophys.*, v.133, p.185, 1989.
- [19] P. Marszalek, J. J. Zielinsky, M. Fikus, and T. Y. Tsong, "Determination of electric parameters of cell membranes by a dielectrophoresis method.," *Biophys. J.*, vol. 59, pp. 982-987, 1991.
- [20] Y. Huang, R. Holzel, R. Pethig, and X. B. Wang, "Differences in AC electrodynamics of viable and non-viable yeast cells determined through combined dielectrophoresis and electrorotation studies.," *Phys. Med. Biol.*, vol. 37, pp. 1449-1517, 1992.
- [21] K. V. I. S. Kaler, J. P. Xie, T. B. Jones, and R. Paul, "Dual-frequency dielectrophoretic levitation of canola protoplasts.," *Biophys. J.*, vol. 63, pp. 58-69, 1992.
- [22] A. V. Sokirko, "The electrorotation of axisymmetrical cell.," *Biol. Mem.*, v. 6, pp. 587-600, 1992.
- [23] P. R. C. Gascoyne, R. Pethig, J. P. H. Burt, and F. F. Becker, "Membrane changes accompanying the induced differentiation of Friend murine erythroleukaemic cells studied by dielectrophoresis," *Biochim. Biophys. Acta.*, vol. 1149, pp. 119-126, 1993.
- [24] T. Muller, L. Kuchler, G. Fuhr, T. Schnelle, and A. Sokirko, "Dielektrische einzelzellspektroskopie an pollern verschiedener waldbaumarten - charakterisierung der pollenvitalitat.," *Silvia Genet.*, vol. 42, pp. 311-322, 1993.
- [25] V. L. Sukhorukov, W. M. Arnold, and U. Zimmerman., "Hypotonically induced changes in the plasma membrane of cultured mammalian cells.," *J. Membr. Biol.*, vol. 132, pp. 27-40, 1993.
- [26] J. Gimsa, T. Muller, T. Schnelle, and G. Fuhr, "Dielectric spectroscopy of single human erythrocytes at physiological ionic strength: dispersion of the cytoplasm.," presented *Biophysics J.*, 1996.
- [27] J. Gimsa, P. Marszalek, U. Lowe, and T. Y. Tsong, "Dielectrophoresis and electrorotation of slime and murine myeloma cells.," *Biophys. J.*, vol. 60, pp. 5-14, 1991.
- [28] H. E. Ayliffe, "Micromachined cellular characterization system for studying the biomechanics of individual cells," presented at IEEE Transducers '97, Chicago, IL, 1997.
- [29] H.E. Ayliffe, R.D. Rabbitt, A.B. Frazier, "Electrical Impedance Spectroscopy using Microchannels with Integrated Metal Electrodes", *IEEE Journal on Microelectromechanical Systems*, (8) 50-57 (1999).
- [30] Sisson, R.M., Giddings, J.C. Nov. 15, 1994. *Effects of solvent composition on polymer retention in TFFF*. *Analytical Chem.*, 66, pp. 4043-4053.
- [31] Caldwell, K.D. 1987. *Field-flow fractionation of biological materials*. Dept. of Bioeng., U. of Utah.
- [32] Lou, J.Z. March 1994. *Studies on the theory and applications of polymer separation by thermal field-flow fractionation*. Ph.D. Thesis, University of Utah, Salt Lake City, Utah. p. 37.
- [33] B.K. Gale, A.B. Frazier, K.D. Caldwell, *Characterization of a  $\mu$ -EFFF system*, MEMS '97, 119-124.
- [34] Hovingh, M.E., Thompson, G.H., Giddings, J.C. Feb. 1970. *Column parameters in TFFF*. *Analytical Chem.*, 42, pp. 195-120
- [35] B.K. Gale, K.D. Caldwell, and A.B. Frazier, "A Micromachined Electrical Field- Flow Fractionation System," *IEEE Tran. on Biomedical Engineering*, vol. 45, no. 12, pp. 1459-1469, 1998.

---

[36] B. K. Gale, K. D. Caldwell, and A. B. Frazier, "Electrical conductivity particle detector for use in biological and chemical micro-analysis systems," in *Proc. SPIE Symposium on Micromachining and Microfabrication: Micro Fluidic Devices and Systems*, 1998, pp. 230-242.

[37] B.K. Gale and A.B. Frazier, " Electrical impedance-spectroscopy particle detector for use in microanalysis systems," in *Proc. SPIE Symposium on Micromachining and Microfabrication: Micro Fluidic Devices and Systems*, 1998.